Sara® Stedy



Sara Stedy and Sara Stedy Compact enable a single caregiver to assist patients or residents perform sit to stand transfers throughout the day.

ARJO MOBILITY GALLERY



© Arjo 2019

B. Ambulatory to some degree but may use a walking frame/crutches for support. Please contact Arjo for further information.

• Promote early mobilisation

Sara Stedy can support early mobility initiatives by encouraging patients to be active during sitting to standing transfers.

• Available in two sizes

Sara Stedy enables a single caregiver to assist sit-to-stand tasks for patients of different sizes and clinical assessments.

• Simplified everyday transfers

Sara Stedy and Sara Stedy Compact are non-powered devices.

• DSDC Dementia Design Accredited Product

The Dementia Services Development Centre at the University of Stirling has reviewed Sara Stedy and awarded it a DSDC Product Accreditation Rating Class 1B.





Key features



| PRODUCT INFORMATION | | |
|--|--|--|
| | Sara Stedy | Sara Stedy Compact |
| Max safe working load | 182 kg/400 lb | 182 kg/400 lb |
| Lifter weight | 29.4 kg/64.82 lb | 28.3 kg/62.5 lb |
| Min total storage height | 1050 mm/41 ⅓ in | 960 mm/37.8 in |
| External width legs closed | 490 mm/19 ⅓ in | 490 mm/19 ⅓ in |
| External width legs fully open | 880 mm/34 ² / ₃ in | 880 mm/34 ² / ₃ in |
| Internal width leg closed | 360 mm/14 ⅓ in | 360 mm/14 ⅓ in |
| Internal width legs fully open | 750 mm/29 ⅓ in | 750 mm/29 ⅓ in |
| Leg clearance (floor to top of legs) | 100 mm/3 % in | 100 mm/3 % in |
| Chassis clearance (floor to bottom of chassis) | 30 mm/1 ⅓ in | 30 mm/1 ⅓ in |
| Lifter total length | 920 mm/36 ⅓ in | 890 mm/35 in |
| Turning radius | 1040 mm/41 in | 1020 mm/40 1/8 in |
| | | |

Low-friction castors, the rear two with brakes


The Dementia Services Development Centre (DSDC) at the University of Stirling has reviewed this product and has rated its design in relation to dementia design principles and its usability in a dementia-inclusive environment. For ease of identification, the suitability of the product is classified by a number with a corresponding explanation of use, within the ratings key provided. The DSDC Accredited Product logo is awarded on the merit of the individual product and colour way. This accreditation does not relate to the manufacture or final application of the product. While every endeavour has been made to ensure the ratings reflect the product's suitability and application at the time of printing, the DSDC cannot be held responsible for the application of the final product, its performance or its interface with other products or finishes. For an environment to be considered 'dementia-inclusive' or 'dementia-friendly', careful consideration must be given to the specification of adjoining finishes, their performance specification (for example light reflectance value (LRV), slip resistance and use of pattern) and their suitability for their intended use.

ACCESSORIES

Wide range of washable, disposable and wipeable slings available

Arjo disinfectants

SPACE REQUIREMENT



- Blue area shows the minimum working area required to enable the caregiver to use the mechanical aids ergonomically from one side.
- Light blue area shows extension of working area required to facilitate activities from either side to provide adequate access for the patient, mechanical aid and caregiver.

